


<b>DI155A-EN</b> (Rev F)	<b>REVLOK® FENESTRATED SCREW SYSTEM</b>	
05/2025  <b>GLOBUS MEDICAL</b> GLOBUS MEDICAL, INC. Valley Forge Business Center 2560 General Armistead Avenue Audubon, PA 19403 USA Customer Service: Phone 1-866-GLOBUS1 (OR) 1-866-456-2871 Fax 1-866-GLOBUS3 (OR) 1-866-456-2873	<b>IMPORTANT INFORMATION ON THE REVLOK® FENESTRATED SCREW SYSTEM</b>	

For symbols glossary, please refer to [www.globusmedical.com/eIFU](http://www.globusmedical.com/eIFU)

ENGLISH

WITHIN THE UNITED STATES ONLY

## IMPORTANT INFORMATION ON REVLOK® FENESTRATED SCREW SYSTEM

### DESCRIPTION

The REVLOK® Fenestrated Screw System consists of monoaxial screws, uniplanar screws, polyaxial screws, dual outer diameter screws, reduction screws, rods and locking caps. Screws and rods are available in a variety of sizes to accommodate individual patient anatomy. REVLOK® implants mate with 5.5mm diameter rods and REVLOK® 6.35 implants mate with 6.35mm diameter rods, and connecting components from the REVERE® Stabilization System. Implant components can be rigidly locked into a variety of configurations for the individual patient and surgical condition. Locking caps are used to connect the screws to the rod.

The most common use of this screw and rod system in the posterior thoracolumbar and sacral spine is two rods, each positioned and attached lateral to the spinous process via pedicle screws.

The most common use of this screw and rod system in the anterior thoracolumbar spine is one rod, positioned and attached to the vertebral bodies via monoaxial screws through an appropriate size staple.

Screws attach to the rods using a locking cap with an inner set screw. The size and number of screws are dependent on the length and location of the rod. Screws are inserted into a pedicle of the thoracolumbar and/or sacral spine. Screws may be used with staples. Cement augmentation is performed through the screw before rod insertion. The length of the rod is dependent on the number and location of the screws.

The rods are composed of titanium alloy, commercially pure titanium, cobalt chromium molybdenum alloy, or stainless steel, as specified in ASTM F136, F1295, F67, F1537 and F138. All other REVLOK® implants are manufactured from titanium alloy or stainless steel, as specified in ASTM F136, F1295, F138, and F67. Due to the risk of galvanic corrosion following implantation, stainless steel implants should not be connected to titanium or titanium alloy implants. The REVLOK® Fenestrated Screws are available with or without hydroxyapatite (HA) coating, as specified in ASTM F1185.

### INDICATIONS

REVLOK® Fenestrated Screw System, when used as a posterior pedicle screw system, is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar and sacral spine: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), degenerative spondylolisthesis with objective evidence of neurologic impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor, pseudoarthrosis and failed previous fusion.

In addition, the REVLOK® Fenestrated Screw System is intended for treatment of severe spondylolisthesis (Grades 3 and 4) of the L5-S1 vertebra in skeletally mature patients receiving fusion by autogenous bone graft, having implants attached to the lumbosacral spine and/or ilium with removal of the implants after attainment of a solid fusion. Levels of pedicle screw fixation for these patients are L3-sacrum/iliac.

When used as an anterolateral thoracolumbar system, the REVLOK® Fenestrated Screw System is intended for anterolateral screw (with or without staple) fixation for the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), spinal stenosis, spondylolisthesis, spinal deformities (i.e. scoliosis, kyphosis, and/or lordosis), fracture or dislocation of the thoracolumbar spine, pseudoarthrosis, tumor resection, and/or failed previous fusion. Levels of screw fixation are T8-L5.

When used for posterior fixation in conjunction with FORTRESS® or FORTRESS® Plus polymethylmethacrylate (PMMA) bone cement the REVLOK® Fenestrated Screw System is intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the thoracic and lumbar spine in whom life expectancy is of insufficient duration to permit achievement of fusion.

REVLOK® Fenestrated screws augmented with FORTRESS® or FORTRESS® Plus bone cement are for use at spinal levels where the structural integrity of the spine is not severely compromised.

### WARNINGS

The safety and effectiveness of pedicle screw spinal systems have been established only for spinal conditions with significant mechanical instability or deformity requiring fusion with instrumentation. These conditions are significant mechanical instability or deformity of the thoracic, lumbar, and sacral spine secondary to degenerative disc disease, degenerative spondylolisthesis with objective evidence of neurologic impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor and failed previous fusion (pseudoarthrosis). The safety and effectiveness of these devices for any other conditions are unknown.

One of the potential risks identified with this system is death. Other potential risks which may require additional surgery, include:

- device component fracture,
- loss of fixation,
- non-union,
- fracture of the vertebrae,
- neurological injury, and
- vascular or visceral injury.

Potential risks when used with bone cement include:

- Hypersensitivity reactions in susceptible persons resulting in anaphylactic response
- Tissue damage, nerve, or circulatory problems caused by cement leakage
- Micromotion of cement against bone surface caused by inadequate fixation

Cement leakage may cause tissue damage, nerve or circulatory problems, and other serious adverse events. These risks may increase with the number of spinal levels where bone cement is utilized, and also with the volume of bone cement used.

Serious adverse events, some with fatal outcome, associated with the use of acrylic bone cements in the spine include myocardial infarction, cardiac arrest, cerebrovascular accident, pulmonary embolism, and cardiac embolism. Although the majority of these adverse events present early within the post-operative period, there have been some reports of diagnoses beyond a year or more after the procedure.

Other reported adverse events for acrylic bone cements intended for use in the spine include leakage of the bone cement beyond the site of its intended application with introduction into the vascular system resulting in embolism of the lung and/or heart or other clinical sequelae.

There is no clinical data regarding the use of bone cement in pregnant or lactating women.

Strict adherence to the surgical technique guide is strongly recommended.

Cement augmentation is not intended for use in screws placed bicortically.

Components of this system should not be used with components of any other system or manufacturer, unless connected through the rod(s).

The components of this system are manufactured from titanium alloy, pure titanium, stainless steel and cobalt chromium-molybdenum alloy. Mixing of stainless steel implant components with different materials is not recommended for metallurgical, mechanical and functional reasons.

If bone cement is seen outside of the vertebral body or in the circulatory system during cement augmentation immediately stop the injection.

### PRECAUTIONS

The implantation of screw, hook and rod systems should be performed only by experienced spinal surgeons with specific training in the use of this system because this is a technically demanding procedure presenting a risk of serious injury to the patient. Preoperative planning and patient anatomy should be considered when selecting screw diameter and length, and hook size.

The REVLOK® Fenestrated 5.5mm and 6.35mm implants are intended for use with a REVERE® and REVOLVE® 5.5mm rod and REVERE® 6.35mm rod.

Surgical implants are SINGLE USE ONLY and must never be reused. An explanted implant must never be reimplanted. Even though the device appears undamaged, it may have small defects and internal stress patterns which could lead to breakage.

For optimal implant performance, when using the REVLOK® Fenestrated Screws System, the physicians/surgeon should consider the levels of implantation, patient weight, patient activity level, other patient conditions, etc., which may impact on the performance of this system.

When performing cement augmentation, confirm that the pedicle length is sufficient for the most posterior screw fenestration to be located within the vertebral body.

### ATTENTION

See Warnings, Precautions and Potential Adverse Events sections of the insert entitled "Suggestions Concerning Orthopaedic Metallic Internal Fixation Devices" for a complete list of potential risks.

### CONTRAINDICATIONS

Certain degenerative diseases or underlying physiological conditions such as diabetes or rheumatoid arthritis may alter the healing process, thereby increasing the risk of implant breakage.

Mental or physical impairment which compromises a patient's ability to comply with necessary limitations or precautions may place that patient at a particular risk during postoperative rehabilitation.

Factors such as the patient's weight, activity level, and adherence to weight bearing or load bearing instructions have an effect on the stresses to which the implant is subjected.

Use of these implants is contraindicated when used with bone cement in patients with the following conditions:

1. Poor visibility under fluoroscopy
2. Patients with thrombophilia
3. Patients with severe cardiac and/or pulmonary insufficiency
4. Patients with known sensitivity to any of the components of bone cement
5. Any patient with a T-score of > -2.5

#### MRI SAFETY INFORMATION

The REVLOK® Fenestrated Screws have not been evaluated for safety and compatibility in the MR environment. These devices have not been tested for heating, migration, or image artifact in the MR environment. The safety of REVLOK® in the MR environment is unknown. Scanning a patient who has these devices may result in patient injury.

#### PACKAGING

These implants and instruments may be supplied pre-packaged and sterile, using gamma irradiation. The integrity of the sterile packaging should be checked to ensure that sterility of the contents is not compromised. Packaging should be carefully checked for completeness and all components should be carefully checked to ensure that there is no damage prior to use. Damaged packages or products should not be used, and should be returned to Globus Medical. During surgery, after the correct size has been determined, remove the products from the packaging using aseptic technique.

The instrument sets are provided nonsterile and are steam sterilized prior to use, as described in the STERILIZATION section below. Following use or exposure to soil, instruments must be cleaned, as described in the CLEANING section below.

#### HANDLING

All instruments and implants should be treated with care. Improper use or handling may lead to damage and/or possible malfunction. Products should be checked to ensure that they are in working order prior to surgery. All products should be inspected prior to use to ensure that there is no unacceptable deterioration such as corrosion, discoloration, pitting, cracked seals, etc. Non-working or damaged instruments should not be used, and should be returned to Globus Medical.

#### CLEANING

All instruments that can be disassembled must be disassembled for cleaning. All handles must be detached. Instruments may be reassembled following sterilization. The instruments should be cleaned using neutral cleaners before sterilization and introduction into a sterile surgical field or (if applicable) return of the product to Globus Medical.

Cleaning and disinfecting of instruments can be performed with aldehyde-free solvents at higher temperatures. Cleaning and decontamination must include the use of neutral cleaners followed by a deionized water rinse. Note: certain cleaning solutions such as those containing formalin, glutaraldehyde, bleach and/or other alkaline cleaners may damage some devices, particularly instruments; these solutions should not be used.

The following cleaning methods should be observed when cleaning instruments after use or exposure to soil, and prior to sterilization:

1. Immediately following use, ensure that the instruments are wiped down to remove all visible soil and kept from drying by submerging or covering with a wet towel.
2. Disassemble all instruments that can be disassembled.
3. Rinse the instruments under running tap water to remove all visible soil. Flush the lumens a minimum of 3 times, until the lumens flush clean.
4. Prepare Enzol® (or a similar enzymatic detergent) per manufacturer's recommendations.
5. Immerse the instruments in the detergent and allow them to soak for a minimum of 2 minutes.
6. Use a soft bristled brush to thoroughly clean the instruments. Use a pipe cleaner for any lumens. Pay close attention to hard to reach areas.
7. Using a sterile syringe, draw up the enzymatic detergent solution. Flush any lumens and hard to reach areas until no soil is seen exiting the area.
8. Remove the instruments from the detergent and rinse them in running warm tap water.
9. Prepare Enzol® (or a similar enzymatic detergent) per manufacturer's recommendations in an ultrasonic cleaner.
10. Completely immerse the instruments in the ultrasonic cleaner and ensure detergent is in lumens by flushing the lumens. Sonicate for a minimum of 3 minutes.
11. Remove the instruments from the detergent and rinse them in running deionized water or reverse osmosis water for a minimum of 2 minutes.
12. Dry instruments using a clean soft cloth and filtered pressurized air.
13. Visually inspect each instrument for visible soil. If visible soil is present, then repeat cleaning process starting with Step 3.

#### CONTACT INFORMATION

Globus Medical may be contacted at 1-866-GLOBUS1 (456-2871). A surgical technique manual may be obtained by contacting Globus Medical.

#### STERILIZATION

These implants and instruments may be available sterile or nonsterile. HA-coated implants are only available sterile.

Sterile implants and instruments are sterilized by gamma radiation, validated to ensure a Sterility Assurance Level (SAL) of  $10^{-6}$ . Sterile products are packaged in a heat sealed, double foil pouch. The expiration date is provided in the package label. These products are considered sterile unless the packaging has been opened or damaged. Sterile implants and instruments that become nonsterile or have expired packaging are considered nonsterile and may be sterilized according to instructions for nonsterile implants and instruments below, with the exception of HA-coated implants, which cannot be resterilized and should be disposed of according to hospital protocol. Sterile implants meet pyrogen limit specifications.

Nonsterile implants and instruments have been validated to ensure an SAL of  $10^{-6}$ . The use of an FDA-cleared wrap is recommended, per the Association for the Advancement of Medical Instrumentation (AAMI) ST79, *Comprehensive Guide to Steam Sterilization and Sterility Assurance in Health Care Facilities*. It is the end user's responsibility to use only sterilizers and accessories (such as sterilization wraps, sterilization pouches, chemical indicators, biological indicators, and sterilization cassettes) that have been cleared by the FDA for the selected sterilization cycle specifications (time and temperature).

When using a rigid sterilization container, the following must be taken into consideration for proper sterilization of Globus devices and loaded graphic cases:

- Recommended sterilization parameters are listed in the table below.
- Only FDA-cleared rigid sterilization containers for use with pre-vacuum steam sterilization may be used.
- When selecting a rigid sterilization container, it must have a minimum filter area of 176 in<sup>2</sup> total, or a minimum of four (4) 7.5in diameter filters.
- No more than one (1) loaded graphic case or its contents can be placed directly into a rigid sterilization container.
- Stand-alone modules/racks or single devices must be placed, without stacking, in a container basket to ensure optimal ventilation.
- The rigid sterilization container manufacturer's instructions for use are to be followed; if questions arise, contact the manufacturer of the specific container for guidance.
- Refer to AAMI ST79 for additional information concerning the use of rigid sterilization containers.

For implants and instruments provided NONSTERILE, sterilization is recommended (wrapped or containerized) as follows:

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Pre-vacuum	132°C (270°F)	4 minutes	30 minutes
Steam	Pre-vacuum	134°C (273°F)	3 minutes	30 minutes

*These parameters are validated to sterilize only this device. If other products are added to the sterilizer, the recommended parameters are not valid and new cycle parameters must be established by the user. The sterilizer must be properly installed, maintained, and calibrated. Ongoing testing must be performed to confirm inactivation of all forms of viable microorganisms.*

**CAUTION:** Federal (U.S.A) Law Restricts this Device to Sale by or on the order of a Physician.