

03/2021



IMPORTANT INFORMATION ON THE CREO® STABILIZATION SYSTEM



**GLOBUS MEDICAL, INC.**  
Valley Forge Business Center  
2560 General Armistead Avenue  
Audubon, PA 19403  
USA  
Customer Service:  
Phone 1-866-GLOBUS1 (OR)  
1-866-456-2871  
Fax 1-866-GLOBUS3 (OR)  
1-866-456-2873

**WITHIN THE UNITED STATES ONLY****ENGLISH****IMPORTANT INFORMATION  
ON THE CREO® STABILIZATION SYSTEM****DESCRIPTION**

The CREO® Stabilization System consists of rods, hooks, monoaxial screws, uniplanar screws, polyaxial screws, reduction screws, fenestrated screws, awl tip screws, locking caps, t-connectors, head offset connectors, trans-iliac connectors, staples, and associated manual surgical instruments. Implants are available in a variety of sizes to accommodate individual patient anatomy. CREO® implants mate with 4.75mm, 5.5mm, and 6.35mm diameter rods. In addition, CREO® 5.5 Threaded screws and locking caps mate with 6.0mm diameter rods. CREO NXT™ and CREO® Preferred Angle implants mate with 5.5mm and 6.0mm rods. CREO DLX™ implants mate with 6.0 and 6.35mm rods. Implant components can be rigidly locked into a variety of configurations for the individual patient and surgical condition. Polyaxial screws, hooks, and t-connectors are intended for posterior use only. Staples are intended for anterior use only. Rods and monoaxial screws may be used anteriorly or posteriorly. Locking caps are used to connect screws or hooks to the rod and trans iliac connectors.

The most common use of this screw, hook, and rod system in the posterior thoracolumbar and sacral spine is two rods, each positioned and attached lateral to the spinous process via pedicle screws and/or lamina, pedicle or transverse process hooks.

The most common use of this screw, hook, and rod system in the anterior thoracolumbar spine is one rod, positioned and attached to the vertebral bodies via monoaxial screws through an appropriate size staple.

Screws and hooks attach to the rods using a locking cap with an inner set screw, or a threaded locking cap. The size and number of screws are dependent on the length and location of the rod. Screws are inserted into a pedicle of the thoracolumbar and/or sacral spine. Screws may be used with a staple. The type and number of hooks are also dependent on the location in the spine needing correction and/or stabilization. Hooks are attached to the laminae, pedicles, or transverse process of the posterior spine.

T-connectors are modular components designed to connect the two rods of a construct and act as a structural cross member. The rod-clamping set screws secure the t-connectors to the rods. Additional set screws secure the adjustable cross members at the desired length. Additional connectors may be used to connect two rods, and are also secured using set screws.

CREO® implants are composed of titanium alloy, cobalt chromium molybdenum alloy, or stainless steel, as specified in ASTM F136, F1295, F1472, F1537 and F138. Rods are also available in commercially pure titanium, as specified in ASTM F67. Screws are also available with hydroxyapatite (HA) coating per ASTM F1185. Due to the risk of galvanic corrosion following implantation, stainless steel implants should not be connected to titanium, titanium alloy, or cobalt chromium-molybdenum alloy implants.

The CREO® System includes manual surgical instruments manufactured from stainless steel, as specified in ASTM F899. Navigation Instruments are nonsterile, reusable instruments that can be operated manually or under power using a power drill such as POWEREASE™, that are intended to be used with the Medtronic StealthStation® System.

CREO ONE™ Robotic Screws are used with ExcelsiusGPS®, Medtronic StealthStation®, or without navigation or guidance assistance. CREO ONE™ Robotic Screws should not be used with any other third-party robotic or navigation system.

**INDICATIONS**

The CREO® Stabilization System implants are non-cervical spinal fixation devices intended for posterior pedicle screw fixation (T1-S2/ilium), posterior hook fixation (T1-L5), or anterolateral fixation (T8-L5). Pedicle screw fixation is indicated for skeletally mature patients (including small stature) and for pediatric patients. These devices are indicated as an adjunct to fusion for the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (i.e., fracture or dislocation), deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis, Scheuermann's Disease), tumor, stenosis, and failed previous fusion (pseudoarthrosis). When used as an adjunct to fusion, the CREO® Stabilization System is intended to be used with autograft and/or allograft.

In addition, the CREO® Stabilization System is intended for treatment of severe spondylolisthesis (Grades 3 and 4) of the L5-S1 vertebra in skeletally mature patients receiving fusion by autogenous bone graft, having implants attached to the lumbosacral spine and/or ilium with removal of the implants after attainment of a solid fusion. Levels of pedicle screw fixation for these patients are L3-sacrum/ilium.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the CREO® Stabilization System implants are indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis. The CREO® Stabilization System is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

In order to achieve additional levels of fixation, the CREO® Stabilization System rods may be connected to the REVERE® Stabilization System (4.5mm, 5.5mm, or 6.35mm rod) or ELLIPSE® Occipito-Cervico-Thoracic Spinal System (3.5mm rod) using corresponding connectors. Refer to the REVERE®, or ELLIPSE® system package insert for instructions and indications of use.

In-Line Connector Growing Rods are indicated in patients under 10 years of age with potential for additional spine growth who require surgical treatment to obtain and maintain correction of severe, progressive, life-threatening, early onset spinal deformities associated with thoracic insufficiency, including early onset scoliosis, as part of a growing rod construct.

Globus Navigation Instruments are intended to be used during the preparation and placement of CREO® screws during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. These instruments are designed for use with the Medtronic StealthStation® System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks of the anatomy.

When used for posterior fixation in conjunction with FORTRESS® or FORTRESS-Plus® bone cement, the CREO® Fenestrated Screw System is intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the thoracic and lumbar spine in whom life expectancy is of insufficient duration to permit achievement of fusion. CREO® Fenestrated screws augmented with FORTRESS™ and FORTRESS-Plus™ bone cements are for use at spinal levels where the structural integrity of the spine is not severely compromised.

**WARNINGS**

The safety and effectiveness of pedicle screw spinal systems have been established only for spinal conditions with significant mechanical instability or deformity requiring fusion with instrumentation. These conditions are significant mechanical instability or deformity of the thoracic, lumbar, and sacral spine secondary to degenerative disc disease, degenerative spondylolisthesis with objective evidence of neurologic impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor, and failed previous fusion (pseudoarthrosis). The safety and effectiveness of these devices for any other conditions are unknown.

One of the potential risks identified with this system is death. Other potential risks which may require additional surgery, include:

- device component fracture,
- loss of fixation,
- non-union,
- fracture of the vertebrae,
- neurological injury, and
- vascular or visceral injury.

Potential risks when used with bone cement include:

- Hypersensitivity reactions in susceptible persons resulting in anaphylactic response
- Tissue damage, nerve, or circulatory problems caused by cement leakage
- Micromotion of cement against bone surface caused by inadequate fixation

Cement leakage may cause tissue damage, nerve or circulatory problems, and other serious adverse events. These risks may increase with the number of spinal levels where bone cement is utilized, and also with the volume of bone cement used.

Serious adverse events, some with fatal outcome, associated with the use of acrylic bone cements in the spine include myocardial infarction, cardiac arrest, cerebrovascular accident, pulmonary embolism, and cardiac embolism. Although the majority of these adverse events present early within the post-operative period, there have been some reports of diagnoses beyond a year or more after the procedure.

Other reported adverse events for acrylic bone cements intended for use in the spine include leakage of the bone cement beyond the site of its intended application with introduction into the vascular system resulting in embolism of the lung and/or heart or other clinical sequelae.

If bone cement is seen outside of the vertebral body or in the circulatory system during cement augmentation immediately stop the injection.

There is no clinical data regarding the use of bone cement in pregnant or lactating women.

Strict adherence to the surgical technique guide is strongly recommended.

Cement augmentation is not intended for use in screws placed bicortically.

Components of this system should not be used with components of any other manufacturer.

The components of this system are manufactured from titanium alloy, pure titanium, stainless steel and cobalt chromium-molybdenum alloy. Mixing of stainless steel implant components with different materials is not recommended for metallurgical, mechanical and functional reasons.

**ADDITIONAL WARNINGS FOR PEDIATRIC PATIENTS**

The use of pedicle screw fixation in the pediatric population may present additional risks when patients are of smaller stature and skeletally immature. Pediatric patients may have smaller spinal structures (pedicle diameter or length) that may preclude the use of pedicle screws or increase the risk of pedicle screw malpositioning and neurological or vascular injury. Patients not skeletally mature that undergo spinal fusion procedures may have reduced longitudinal spinal growth, or may be at risk for rotational spinal deformities ("crankshaft phenomenon") due to continued differential growth of the anterior spine.

Pediatric patients may be at increased risk for device-related injury because of their smaller stature.

**PRECAUTIONS**

The implantation of screw, hook and rod systems should be performed only by experienced spinal surgeons with specific training in the use of this system because this is a technically demanding procedure presenting a risk of serious injury to the patient. Preoperative planning and patient anatomy should be considered when selecting screw diameter and length, and hook size.

The CREO® Stabilization System includes 4.75 implants intended for use with a 4.75mm rod, 5.5 implants intended for use with a 5.5mm rod, and 6.35 implants intended for use with a 6.35mm rod. CREO® 5.5 Threaded screws and locking caps are also intended for use with a 6.0mm rod. CREO NXT™ and CREO® Preferred Angle implants are intended for use with 5.5mm and 6.0mm rods and CREO DLX™ implants are intended for use with 6.0mm and 6.35mm rods.

Surgical implants are SINGLE USE ONLY and must never be reused. An explanted implant must never be reimplanted. Even though the device appears undamaged, it may have small defects and internal stress patterns which could lead to breakage.

Based on fatigue testing results, when using the CREO® Stabilization System, the surgeon should consider the levels of implantation, patient weight, patient activity level, other patient conditions, etc., which may impact on the performance of this system.

When performing cement augmentation, confirm that the pedicle length is sufficient for the most posterior screw fenestration to be located within the vertebral body.

## ADDITIONAL PRECAUTIONS FOR PEDIATRIC PATIENTS

The implanting surgeon should consider carefully the size and type of implants most suitable for the pediatric patient's age, size, weight and skeletal maturity.

Since pediatric patients may have additional growth potential following implant surgery, the likelihood of a subsequent removal and/or revision surgery is greater than in adult patients.

## MRI SAFETY INFORMATION

CREO® has not been evaluated for safety and compatibility in the MR environment. CREO® has not been tested for heating, migration, or image artifact in the MR environment. The safety of CREO® in the MR environment is unknown. Scanning a patient who has these devices may result in patient injury.

## CONTRAINDICATIONS

Certain degenerative diseases or underlying physiological conditions such as diabetes or rheumatoid arthritis may alter the healing process, thereby increasing the risk of implant breakage.

Mental or physical impairment which compromises a patient's ability to comply with necessary limitations or precautions may place that patient at a particular risk during postoperative rehabilitation.

Factors such as the patient's weight, activity level, and adherence to weight bearing or load bearing instructions have an effect on the stresses to which the implant is subjected.

Use of these implants is contraindicated in patients with the following conditions:

1. Active systemic infection, infection or inflammation localized to the site of the proposed implantation, or when the patient has demonstrated allergy or foreign body sensitivity to any of the implant materials.
2. Prior fusion at the level(s) to be treated.
3. Severe osteoporosis, which may prevent adequate fixation.
4. Conditions that may place excessive stresses on bone and implants, such as severe obesity or degenerative diseases, are relative contraindications. The decision whether to use these devices in such conditions must be made by the physician taking into account the risks versus the benefits to the patient.
5. Patients whose activity, mental capacity, mental illness, alcoholism, drug abuse, occupation, or lifestyle may interfere with their ability to follow postoperative restrictions and who may place undue stresses on the implant during bony healing and may be at a higher risk of implant failure.
6. Any patient not willing to cooperate with postoperative instruction.
7. Any condition not described in the indications for use.
8. Fever or leukocytosis.
9. Pregnancy.
10. Any other condition which would preclude the potential benefit of spinal implant surgery, such as the presence of tumors or congenital abnormalities, fracture local to the operating site, elevation of sedimentation rate unexplained by other diseases, elevation of the white blood count (WBC), or a marked left shift in the WBC differential count.
11. Patients with a known hereditary or acquired bone friability or calcification problem should not be considered for this type of surgery.
12. Any case where the implant components selected for use would be too large or too small to achieve a successful result.
13. Any case that requires the mixing of metals from two different components or systems.
14. Any patient having inadequate tissue coverage at the operative site or inadequate bone stock or quality.
15. Any patient in which implant utilization would interfere with anatomical structures or expected physiological performance.

Use of these implants is contraindicated when used with bone cement in patients with the following conditions:

1. Poor visibility under fluoroscopy
2. Patients with thrombophilia
3. Patients with severe cardiac and/or pulmonary insufficiency
4. Patients with known sensitivity to any of the components of bone cement
5. Any patient with a T-score of > -2.5

## PACKAGING

These implants and instruments may be supplied pre-packaged and sterile, using gamma irradiation. The integrity of the sterile packaging should be checked to ensure that sterility of the contents is not compromised. Packaging should be carefully checked for completeness and all components should be carefully checked to ensure that there is no damage prior to use. Damaged packages or products should not be used, and should be returned to Globus Medical. During surgery, after the correct size has been determined, remove the products from the packaging using aseptic technique.

The instrument sets are provided nonsterile and are steam sterilized prior to use, as described in the STERILIZATION section below. Following use or exposure to soil, instruments must be cleaned, as described in the CLEANING section below.

## HANDLING AND USE

All instruments and implants should be treated with care. Improper use or handling may lead to damage and/or possible malfunction. Products should be checked to ensure that they are in working order prior to surgery. All products should be inspected prior to use to ensure that there is no unacceptable deterioration such as corrosion (i.e. rust, pitting), discoloration, excessive scratches, notches, debris, residue, flaking, wear, cracks, cracked seals, etc. Non-working or damaged instruments should not be used, and should be returned to Globus Medical.

Any implant that has not been used, but has become soiled, should be handled according to hospital protocol. Any implant with evidence of damage, residue, debris, or other defects should not be used, and should be returned to Globus Medical.

## CLEANING

All instruments that can be disassembled must be disassembled for cleaning. All handles must be detached. Instruments may be reassembled following sterilization. The instruments should be cleaned using neutral cleaners before sterilization and introduction into a sterile surgical field or (if applicable) return of the product to Globus Medical.

Cleaning and disinfecting of instruments can be performed with aldehyde-free solvents at higher temperatures. Cleaning and decontamination must include the use of neutral cleaners followed by a deionized water rinse. Note: certain cleaning solutions such as those containing formalin, glutaraldehyde, bleach and/or other alkaline cleaners may damage some devices, particularly instruments; these solutions should not be used.

The following cleaning methods should be observed when cleaning instruments after use or exposure to soil, and prior to sterilization:

1. Immediately following use, ensure that the instruments are wiped down to remove all visible soil and kept from drying by submerging or covering with a wet towel.
2. Disassemble all instruments that can be disassembled.
3. Rinse the instruments under running tap water to remove all visible soil. Flush the lumens a minimum of 3 times, until the lumens flush clean.
4. Prepare Enzo® (or a similar enzymatic detergent) per manufacturer's recommendations.
5. Immerse the instruments in the detergent and allow them to soak for a minimum of 2 minutes.
6. Use a soft bristled brush to thoroughly clean the instruments. Use a pipe cleaner for any lumens. Pay close attention to hard to reach areas.
7. Using a sterile syringe, draw up the enzymatic detergent solution. Flush any lumens and hard to reach areas until no soil is seen exiting the area.
8. Remove the instruments from the detergent and rinse them in running warm tap water.
9. Prepare Enzo® (or a similar enzymatic detergent) per manufacturer's recommendations in an ultrasonic cleaner.
10. Completely immerse the instruments in the ultrasonic cleaner and ensure detergent is in lumens by flushing the lumens. Sonicate for a minimum of 3 minutes.
11. Remove the instruments from the detergent and rinse them in running deionized water or reverse osmosis water for a minimum of 2 minutes.
12. Dry instruments using a clean soft cloth and filtered pressurized air.
13. Visually inspect each instrument for visible soil. If visible soil is present, then repeat cleaning process starting with Step 3.

## CONTACT INFORMATION

Globus Medical may be contacted at 1-866-GLOBUS1 (456-2871). A surgical technique manual may be obtained by contacting Globus Medical.

## STERILIZATION

These implants and instruments may be available sterile or nonsterile. HA-coated implants are only available sterile.

Sterile implants and instruments are sterilized by gamma radiation, validated to ensure a Sterility Assurance Level (SAL) of 10<sup>-6</sup>. Sterile products are packaged in a heat sealed double pouch or container/pouch. The expiration date is provided in the package label. These products are considered sterile unless the packaging has been opened or damaged.

Sterile implants and instruments that become nonsterile or have expired packaging are considered nonsterile and may be sterilized according to instructions for nonsterile implants and instruments below, with the exception of HA-coated implants, which cannot be re-sterilized and should be disposed of according to hospital protocol. Sterile implants meet pyrogen limit specifications.

Nonsterile implants and instruments have been validated to ensure an SAL of 10<sup>-6</sup>. The use of an FDA-cleared wrap is recommended, per the Association for the Advancement of Medical Instrumentation (AAMI) ST79, *Comprehensive Guide to Steam Sterilization and Sterility Assurance in Health Care Facilities*. It is the end user's responsibility to use only sterilizers and accessories (such as sterilization wraps, sterilization pouches, chemical indicators, biological indicators, and sterilization cassettes) that have been cleared by the FDA for the selected sterilization cycle specifications (time and temperature).

When using a rigid sterilization container, the following must be taken into consideration for proper sterilization of Globus devices and loaded graphic cases:










- Recommended sterilization parameters are listed in the table below.
- Only FDA-cleared rigid sterilization containers for use with pre-vacuum steam sterilization may be used.
- When selecting a rigid sterilization container, it must have a minimum filter area of 176 in<sup>2</sup> total, or a minimum of four (4) 7.5in diameter filters.
- No more than one (1) loaded graphic case or its contents can be placed directly into a rigid sterilization container.
- Stand-alone modules/racks or single devices must be placed, without stacking, in a container basket to ensure optimal ventilation.
- The rigid sterilization container manufacturer's instructions for use are to be followed; if questions arise, contact the manufacturer of the specific container for guidance.
- Refer to AAMI ST79 for additional information concerning the use of rigid sterilization containers.

For implants and instruments provided NONSTERILE, sterilization is recommended (wrapped or containerized) as follows:

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Pre-vacuum	132°C (270°F)	4 minutes	30 minutes
Steam	Pre-vacuum	134°C (273°F)	3 minutes	30 minutes

*These parameters are validated to sterilize only this device. If other products are added to the sterilizer, the recommended parameters are not valid and new cycle parameters must be established by the user. The sterilizer must be properly installed, maintained, and calibrated. Ongoing testing must be performed to confirm inactivation of all forms of viable microorganisms.*

**CAUTION:** Federal Law (USA) Restricts this Device to Sale by or on the order of a Physician.

SYMBOL TRANSLATION			
	CATALOGUE NUMBER		STERILIZED BY IRRADIATION
	LOT NUMBER		AUTHORISED REPRESENTATIVE IN THE EUROPEAN COMMUNITY
	CAUTION		MANUFACTURER
	SINGLE USE ONLY		USE BY (YYYY-MM-DD)
	QUANTITY		