

DI105A-EN (Rev K)	PROTEX® CT OCCIPITO-CERVICO-THORACIC SPINAL SYSTEM
05/2025  GLOBUS MEDICAL GLOBUS MEDICAL, INC. Valley Forge Business Center 2560 General Armistead Avenue Audubon, PA 19403 USA Customer Service: Phone 1-866-GLOBUS1 (OR) 1-866-456-2871 Fax 1-866-GLOBUS3 (OR) 1-866-456-2873	IMPORTANT INFORMATION ON THE PROTEX® CT OCCIPITO-CERVICO-THORACIC SPINAL SYSTEM

For symbols glossary, please refer to www.globusmedical.com/eIFU

ENGLISH

WITHIN THE UNITED STATES ONLY

IMPORTANT INFORMATION ON THE PROTEX® CT OCCIPITO-CERVICO-THORACIC SPINAL SYSTEM

DESCRIPTION

The PROTEX® CT Occipito-Cervico-Thoracic Spinal System consists of rods, polyaxial screws, hooks, locking caps, t-connectors, lateral connectors, parallel connectors, and occipital clamps. The implants are composed of titanium alloy (per ASTM F136, F1472, or F1295) or stainless steel (per ASTM F138). Due to the risk of galvanic corrosion following implantation, stainless steel implants should not be connected to titanium or titanium alloy implants.

The PROTEX® System includes manual surgical instruments manufactured from stainless steel, as specified in ASTM F899. Navigation Instruments are nonsterile, reusable instruments that can be operated manually or under power using a power drill such as POWEREASE™, that are intended to be used with the Medtronic StealthStation® System.

INDICATIONS

The PROTEX® CT Occipito-Cervico-Thoracic Spinal System implants are intended to provide immobilization and stabilization of spinal segments as an adjunct to fusion for the following acute and chronic instabilities of the craniocervical junction, the cervical spine (C1-C7) and the thoracic spine (T1-T3): traumatic spinal fractures and/or traumatic dislocations; instability or deformity; failed previous fusions (e.g. pseudoarthrosis); tumors involving the cervical/thoracic spine; and degenerative disease, including intractable radiculopathy and/or myelopathy, neck and/or arm pain of discogenic origin as confirmed by radiographic studies, and degenerative disease of the facets with instability. These implants are also intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the cervical spine in whom life expectancy is of insufficient duration to permit achievement of fusion.

In order to achieve additional levels of fixation, rods may be connected to occipital cervical thoracic or thoracolumbar stabilization systems ranging in diameter from 3.2mm to 6.5mm, using corresponding connectors.

Globus Navigation Instruments are intended to be used during the preparation and placement of PROTEX® CT screws during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. These instruments are designed for use with the Medtronic StealthStation® System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks of the anatomy.

CONTRAINDICATIONS

Certain degenerative diseases or underlying physiological conditions such as diabetes or rheumatoid arthritis may alter the healing process, thereby increasing the risk of implant breakage.

Mental or physical impairment which compromises a patient's ability to comply with necessary limitations or precautions may place that patient at a particular risk during postoperative rehabilitation.

Factors such as the patient's weight, activity level, and adherence to weight bearing or load bearing instructions have an effect on the stresses to which the implant is subjected.

WARNINGS

The safety and effectiveness of pedicle screw spinal systems have been established only for spinal conditions with significant mechanical instability or deformity requiring fusion with instrumentation. These conditions are significant mechanical instability or deformity of the thoracic spine secondary to degenerative spondylolisthesis with objective evidence of neurological impairment, fracture, dislocation, spinal tumor, and failed previous fusion

(pseudoarthrosis). The safety and effectiveness of these devices for any other conditions are unknown.

Possible adverse effects which may occur and may require additional surgery include: failed fusion or pseudarthrosis leading to implant breakage; allergic reaction to implant materials; device fracture or failure; device migration or loosening; loss of fixation; vertebral fracture; decrease in bone density; pain, discomfort, or abnormal sensations due to the presence of the device; injury to nerves, vessels, and organs; venous thrombosis, lung embolism and cardiac arrest; and death.

The components of this system are manufactured from titanium alloy or stainless steel. Dissimilar metals in contact with each other can accelerate the corrosion process due to galvanic corrosion effects. Mixing of implant components with different materials is not recommended, for metallurgical, mechanical and functional reasons. Components of this system should not be used with components of any other system or manufacturer, unless specifically stated.

These warnings do not include all adverse effects which could occur with surgery in general, but are important considerations particular to orthopedic implants. General surgical risks should be explained to the patient prior to surgery.

PRECAUTIONS

The implantation of posterior screw spinal systems should be performed only by experienced spinal surgeons with specific training in the use of this posterior screw spinal system because this is a technically demanding procedure presenting a risk of serious injury to the patient.

Preoperative planning for implant of cervical posterior screw implants should include review of radiographs, CT and/or MRI imaging to evaluate the patient's anatomy, transverse foramen and the course of the vertebral artery. If any findings would compromise the placement of posterior screws, other surgical methods should be considered. In addition, use of intraoperative imaging should be considered to guide and/or verify device placement, as necessary.

The implants are for single use only. Surgical implants must never be reused. An explanted metal implant must never be reimplanted. Even though the device appears undamaged, it may have small defects and internal stress patterns which could lead to breakage.

Correct handling of the implant is extremely important. Contouring of metal implants should be avoided where possible. If contouring is necessary, or allowed by design, the surgeon should avoid sharp bends, reverse bends, or bending the device at a screw hole. The operating surgeon should avoid any notching or scratching of the device when contouring it. These factors may produce internal stresses which may become the focal point for eventual breakage of the implant.

Metallic implants can loosen, fracture, corrode, migrate, cause pain, or stress shield bone even after a fracture has healed, particularly in young, active patients. While the surgeon must have the final decision on implant removal, we recommend that whenever possible and practical for the individual patient, fixation devices should be removed once their service as an aid to healing is accomplished. Implant removal should be followed by adequate postoperative management.

MRI SAFETY INFORMATION

These devices have not been evaluated for safety and compatibility in the MR environment. It has not been tested for heating, migration, or image artifact in the MR environment. The safety of devices in the MR environment is unknown. Scanning a patient who has this device may result in patient injury.

PACKAGING

These implants and instruments may be supplied pre-packaged and sterile, using gamma irradiation. The integrity of the sterile packaging should be checked to ensure that sterility of the contents is not compromised. Packaging should be carefully checked for completeness and all components should be carefully checked to ensure that there is no damage prior to use. Damaged packages or products should not be used, and should be returned to Globus Medical. During surgery, after the correct size has been determined, remove the products from the packaging using aseptic technique.

The instrument sets are provided nonsterile and are steam sterilized prior to use, as described in the STERILIZATION section below. Following use or exposure to soil, instruments and instrument trays and cases must be cleaned, as described in the CLEANING section below.

HANDLING

All implants, instruments, and instrument trays and cases should be treated with care. Improper use or handling may lead to damage and/or possible malfunction. Products should be checked to ensure that they are in working order prior to surgery. All products should be inspected prior to use to ensure that there is no unacceptable deterioration such as corrosion, discoloration, pitting, cracked seals, etc. Non-working or damaged products should not be used, and should be returned to Globus Medical.

CLEANING

Instruments should be cleaned separately from instrument trays and cases. Lids should be removed from cases for the cleaning process, if applicable. All instruments that can be disassembled must be disassembled for cleaning. All handles must be detached. Instruments may be reassembled following sterilization. The products should be cleaned using neutral cleaners before sterilization and introduction into a sterile surgical field or (if applicable) return of the product to Globus Medical.

Cleaning and disinfecting can be performed with aldehyde-free solvents at higher temperatures. Cleaning and decontamination must include the use of neutral cleaners followed by a deionized water rinse. Note: certain cleaning solutions such as those

containing formalin, glutaraldehyde, bleach and/or other alkaline cleaners may damage some devices, particularly instruments; these solutions should not be used.

The following cleaning methods should be observed when cleaning instruments and instrument trays and cases after use or exposure to soil, and prior to sterilization:

1. Immediately following use, ensure that the instruments are wiped down to remove all visible soil and kept from drying by submerging or covering with a wet towel.
2. Disassemble all instruments that can be disassembled.
3. Rinse the instruments under running tap water to remove all visible soil. Flush the lumens a minimum of 3 times, until the lumens flush clean.
4. Prepare Enzol® (or a similar enzymatic detergent) per manufacturer's recommendations.
5. Immerse the instruments in the detergent and allow them to soak for a minimum of 2 minutes.
6. Use a soft bristled brush to thoroughly clean the instruments. Use a pipe cleaner for any lumens. Pay close attention to hard to reach areas.
7. Using a sterile syringe, draw up the enzymatic detergent solution. Flush any lumens and hard to reach areas until no soil is seen exiting the area.
8. Remove the instruments from the detergent and rinse them in running warm tap water.
9. Prepare Enzol® (or a similar enzymatic detergent) per manufacturer's recommendations in an ultrasonic cleaner.
10. Completely immerse the instruments in the ultrasonic cleaner and ensure detergent is in lumens by flushing the lumens. Sonicate for a minimum of 3 minutes.
11. Remove the instruments from the detergent and rinse them in running deionized water or reverse osmosis water for a minimum of 2 minutes.
12. Dry instruments using a clean soft cloth and filtered pressurized air.
13. Visually inspect each instrument for visible soil. If visible soil is present, then repeat cleaning process starting with Step 3.

CONTACT INFORMATION

Globus Medical may be contacted at 1-866-GLOBUS1 (456-2871). A surgical technique manual may be obtained by contacting Globus Medical.

STERILIZATION

These implants and instruments may be available sterile or nonsterile.

Sterile implants and instruments are sterilized by gamma radiation, validated to ensure a Sterility Assurance Level (SAL) of 10⁻⁶. Sterile products are packaged in a heat sealed, double foil pouch. The expiration date is provided on the package label. These products are considered sterile unless the packaging has been opened or damaged. Sterile implants meet pyrogen limit specifications.

Nonsterile implants and instruments have been validated to ensure an SAL of 10⁻⁶. The use of an FDA-cleared wrap is recommended, per the Association for the Advancement of Medical Instrumentation (AAMI) ST79, *Comprehensive Guide to Steam Sterilization and Sterility Assurance in Health Care Facilities*. It is the end user's responsibility to use only sterilizers and accessories (such as sterilization wraps, sterilization pouches, chemical indicators, biological indicators, and sterilization cassettes) that have been cleared by the FDA for the selected sterilization cycle specifications (time and temperature).

When using a rigid sterilization container, the following must be taken into consideration for proper sterilization of Globus devices and loaded graphic cases:

- Recommended sterilization parameters are listed in the table below.
- Only FDA-cleared rigid sterilization containers for use with pre-vacuum steam sterilization may be used.
- When selecting a rigid sterilization container, it must have a minimum filter area of 176 in² total, or a minimum of four (4) 7.5in diameter filters.
- No more than one (1) loaded graphic case or its contents can be placed directly into a rigid sterilization container.
- Stand-alone modules/racks or single devices must be placed, without stacking, in a container basket to ensure optimal ventilation.
- The rigid sterilization container manufacturer's instructions for use are to be followed; if questions arise, contact the manufacturer of the specific container for guidance.
- Refer to AAMI ST79 for additional information concerning the use of rigid sterilization containers.

For implants and instruments provided NONSTERILE, sterilization is recommended (wrapped or containerized) as follows:

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Pre-vacuum	132°C (270°F)	4 minutes	30 minutes
Steam	Pre-vacuum	134°C (273°F)	3 minutes	30 minutes

Do not stack trays during sterilization. These parameters are validated to sterilize only this device. If other products are added to the sterilizer, the recommended parameters are not valid and new cycle parameters must be established by the user. The sterilizer must be properly installed, maintained, and calibrated. Ongoing testing must be performed to confirm inactivation of all forms of viable microorganisms.

CAUTION: Federal (U.S.A.) Law Restricts this Device to Sale by or on the Order of a Physician.