

D1166B-EN (Rev F)	REVERE™ 4.5 STABILIZATION SYSTEM
05/2025  GLOBUS MEDICAL GLOBUS MEDICAL, INC. Valley Forge Business Center 2560 General Armistead Avenue Audubon, PA 19403 USA Customer Service: Phone 1-866-GLOBUS1 (OR) 1-866-456-2871 Fax 1-866-GLOBUS3 (OR) 1-866-456-2873	IMPORTANT INFORMATION ON THE REVERE™ 4.5 STABILIZATION SYSTEM [CE REP]: AJW Technology Consulting GmbH Breite Straße 3 40213 Düsseldorf, Germany [CH REP]: AJW Technology Consulting GmbH Kreuzplatz 2, 8032 Zurich, Switzerland AUSTRALIA SPONSOR: GLOBUS MEDICAL AUSTRALIA PTY LIMITED, Unit 9/5-7 Inglewood Place Baulkham Hills NSW 2153, Australia  0297 

For symbols glossary, please refer to www.globusmedical.com/eIFU

ENGLISH

OUTSIDE THE UNITED STATES ONLY

IMPORTANT INFORMATION ON THE REVERE™ 4.5 STABILIZATION SYSTEM

DESCRIPTION

The REVERE™ 4.5 Stabilization System consists of rods, hooks, monoaxial screws, uniplanar screws, polyaxial screws; reduction screws, locking caps, t-connectors, head offset connectors, trans-iliac connectors, staples, and associated manual surgical instruments. Screws are available in a variety of sizes to accommodate individual patient anatomy. REVERE™ 4.5 implants mate with 4.5mm diameter rods. Implant components can be rigidly locked into a variety of configurations for the individual patient and surgical condition. Polyaxial screws, hooks, and t-connectors are intended for posterior use only. Staples are intended for anterior use only. Rods and monoaxial screws may be used anteriorly or posteriorly. Locking caps are used to connect screws or hooks to the rod and trans iliac connectors.

The most common use of this screw, hook, and rod system in the posterior thoracolumbar and sacral spine is two rods, each positioned and attached lateral to the spinous process via pedicle screws and/or lamina, pedicle or transverse process hooks.

The most common use of this screw, hook, and rod system in the anterior thoracolumbar spine is one rod, positioned and attached to the vertebral bodies via monoaxial screws through an appropriate size staple.

Screws and hooks attach to the rods using a locking cap with an inner set screw. The size and number of screws are dependent on the length and location of the rod. Screws are inserted into a pedicle of the thoracolumbar and/or sacral spine. The type and number of hooks are also dependent on the location in the spine needing correction and/or stabilization. Hooks are attached to the laminae, pedicles, or transverse process of the posterior spine.

T-connectors are modular components designed to connect the two rods of a construct and act as a structural cross member. The rod-clamping set screws secure the t-connectors to the rods. Additional set screws secure the adjustable cross members at the desired length. REVERE™ 4.5 t-connectors may only be used with 4.5mm diameter rods. Additional connectors may be used to connect two rods, and are also secured using set screws. Connectors may also be used to create growing rod constructs in pediatric patients to allow growth until skeletal maturity.

The rods are composed of titanium alloy, commercially pure titanium, cobalt chromium molybdenum alloy, or stainless steel, as specified in ASTM F136, F1295, F1472, F67, F1537 and F138. All other REVERE™ 4.5 implants are composed of titanium alloy or stainless steel, as specified in ASTM F136, F1295 and F138. Due to the risk of galvanic corrosion following implantation, stainless steel implants should not be connected to titanium, titanium alloy, or cobalt chromium-molybdenum alloy implants.

INDICATIONS

The REVERE™ 4.5 Stabilization System implants are non-cervical spinal fixation devices intended for posterior pedicle screw fixation (T1-S2/ilium), posterior hook fixation (T1-L5), or anterolateral fixation (T8-L5). Pedicle screw fixation is indicated for skeletally mature patients (including small stature) and pediatric patients. These devices are indicated as an adjunct to fusion for all of the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (i.e., fracture or dislocation), deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis, Scheuermann's Disease), tumor, stenosis, and failed previous fusion (pseudoarthrosis).

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the REVERE™ 4.5 Stabilization System implants are indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis. The REVERE™ 4.5 Stabilization System is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

In order to achieve additional levels of fixation in skeletally mature patients, the REVERE™ 4.5 Stabilization System rods may be connected to the REVERE™ Stabilization System (5.5mm or 6.35mm rod) or ELLIPSE™ Occipito-Cervico-Thoracic Spinal System (3.5mm rod)

using corresponding connectors. Refer to the REVERE™ or ELLIPSE™ system package insert for instructions and indications of use.

WARNINGS

The safety and effectiveness of pedicle screw spinal systems have been established only for spinal conditions with significant mechanical instability or deformity requiring fusion with instrumentation. These conditions are significant mechanical instability or deformity of the thoracic, lumbar, and sacral spine secondary to degenerative disc disease, degenerative spondylolisthesis with objective evidence of neurologic impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor and failed previous fusion (pseudoarthrosis). The safety and effectiveness of these devices for any other conditions are unknown.

One of the potential risks identified with this system is death. Other potential risks which may require additional surgery, include:

- device component fracture,
- loss of fixation,
- non-union,
- fracture of the vertebrae,
- changes to spinal curvature,
- neurological injury, and
- vascular or visceral injury.

Components of this system should not be used with components of any other manufacturer.

The components of this system are manufactured from titanium alloy, pure titanium, stainless steel and cobalt chromium-molybdenum alloy. Mixing of stainless steel implant components with different materials is not recommended for metallurgical, mechanical and functional reasons.

ADDITIONAL WARNINGS FOR PEDIATRIC PATIENTS

The use of pedicle screw fixation in the pediatric population may present additional risks when patients are of smaller stature and skeletally immature. Pediatric patients may have smaller spinal structures (pedicle diameter or length) that may preclude the use of pedicle screws or increase the risk of pedicle screw malpositioning and neurological or vascular injury. Patients not skeletally mature that undergo spinal fusion procedures may have reduced longitudinal spinal growth, or may be at risk for rotational spinal deformities ("crankshaft phenomenon") due to continued differential growth of the anterior spine.

Pediatric patients may be at increased risk for device-related injury because of their smaller stature.

ADDITIONAL WARNING (FOR BRAZIL ONLY)

The safety and effectiveness of this device has not been established for use as part of a growing rod construct. This device is only intended to be used when definitive fusion is being performed at all instrumented levels.

PRECAUTIONS

The implantation of screw, hook and rod systems should be performed only by experienced spinal surgeons with specific training in the use of this system because this is a technically demanding procedure presenting a risk of serious injury to the patient. Preoperative planning and patient anatomy should be considered when selecting screw diameter and length, and hook size.

The REVERE™ 4.5 Stabilization System includes implants intended for use with a 4.5mm rod.

Surgical implants are SINGLE USE ONLY and must never be reused. An explanted implant must never be reimplanted. Even though the device appears undamaged, it may have small defects and internal stress patterns which could lead to breakage.

The REVERE™ 4.5 Stabilization System has not been evaluated for safety and compatibility in the MR environment. The REVERE™ 4.5 Stabilization System has not been tested for heating or migration in the MR environment.

Based on fatigue testing results, when using the REVERE™ 4.5 Stabilization System, the physicians/surgeon should consider the levels of implantation, patient weight, patient activity level, other patient conditions, etc., which may impact on the performance of this system.

ADDITIONAL PRECAUTIONS FOR PEDIATRIC PATIENTS

The implanting surgeon should consider carefully the size and type of implants most suitable for the pediatric patient's age, size, weight and skeletal maturity.

Since pediatric patients may have additional growth potential following implant surgery, the likelihood of a subsequent removal and/or revision surgery is greater than in adult patients.

CONTRAINDICATIONS

Certain degenerative diseases or underlying physiological conditions such as diabetes or rheumatoid arthritis may alter the healing process, thereby increasing the risk of implant breakage.

Mental or physical impairment which compromises a patient's ability to comply with necessary limitations or precautions may place that patient at a particular risk during postoperative rehabilitation.

Factors such as the patient's weight, activity level, and adherence to weight bearing or load bearing instructions have an effect on the stresses to which the implant is subjected.

CLEANING

All instruments that can be disassembled must be disassembled for cleaning. All handles must be detached. Instruments may be reassembled following sterilization. The instruments should be cleaned using neutral cleaners before sterilization and introduction into a sterile surgical field or (if applicable) return of the product to Globus Medical.

Cleaning and disinfecting of instruments can be performed with aldehyde-free solvents at higher temperatures. Cleaning and decontamination must include the use of neutral cleaners followed by a deionized water rinse. Note: certain cleaning solutions such as those containing formalin, glutaraldehyde, bleach and/or other alkaline cleaners may damage some devices, particularly instruments; these solutions should not be used.

The following cleaning methods should be observed when cleaning instruments after use or exposure to soil, and prior to sterilization:

1. Immediately following use, ensure that the instruments are wiped down to remove all visible soil and kept from drying by submerging or covering with a wet towel.
2. Disassemble all instruments that can be disassembled.
3. Rinse the instruments under running tap water to remove all visible soil. Flush the lumens a minimum of 3 times, until the lumens flush clean.
4. Prepare Enzo[®] (or a similar enzymatic detergent) per manufacturer's recommendations.
5. Immerse the instruments in the detergent and allow them to soak for a minimum of 2 minutes.
6. Use a soft bristled brush to thoroughly clean the instruments. Use a pipe cleaner for any lumens. Pay close attention to hard to reach areas.
7. Using a sterile syringe, draw up the enzymatic detergent solution. Flush any lumens and hard to reach areas until no soil is seen exiting the area.
8. Remove the instruments from the detergent and rinse them in running warm tap water.
9. Prepare Enzo[®] (or a similar enzymatic detergent) per manufacturer's recommendations in an ultrasonic cleaner.
10. Completely immerse the instruments in the ultrasonic cleaner and ensure detergent is in lumens by flushing the lumens. Sonicate for a minimum of 3 minutes.
11. Remove the instruments from the detergent and rinse them in running deionized water or reverse osmosis water for a minimum of 2 minutes.
12. Dry instruments using a clean soft cloth and filtered pressurized air.
13. Visually inspect each instrument for visible soil. If visible soil is present, then repeat cleaning process starting with Step 3.

CONTACT INFORMATION

Globus Medical may be contacted at 1-866-GLOBUS1 (456-2871). A surgical technique manual may be obtained by contacting Globus Medical.

STERILIZATION

These implants and instruments may be available sterile or nonsterile.

Sterile implants and instruments are sterilized by gamma radiation, validated to ensure a Sterility Assurance Level (SAL) of 10^{-6} . Sterile products are packaged in a heat sealed, double foil pouch. The expiration date is provided in the package label. These products are considered sterile unless the packaging has been opened or damaged.

Nonsterile implants and instruments have been validated to ensure an SAL of 10^{-6} . The use of a wrap is recommended, per the Association for the Advancement of Medical Instrumentation (AAMI) ST79, *Comprehensive Guide to Steam Sterilization and Sterility Assurance in Health Care Facilities*. It is the end user's responsibility to use only sterilizers and accessories (such as sterilization wraps, sterilization pouches, chemical indicators, biological indicators, and sterilization cassettes) that are designed for the selected sterilization cycle specifications (time and temperature).

When using a rigid sterilization container, the following must be taken into consideration for proper sterilization of Globus devices and loaded graphic cases:

- Recommended sterilization parameters are listed in the table below.
- Only rigid sterilization containers for use with pre-vacuum steam sterilization may be used.
- When selecting a rigid sterilization container, it must have a minimum filter area of 176 in² total, or a minimum of four (4) 7.5in diameter filters.
- No more than one (1) loaded graphic case or its contents can be placed directly into a rigid sterilization container.
- Stand-alone modules/racks or single devices must be placed, without stacking, in a container basket to ensure optimal ventilation.
- The rigid sterilization container manufacturer's instructions for use are to be followed; if questions arise, contact the manufacturer of the specific container for guidance.
- Refer to AAMI ST79 for additional information concerning the use of rigid sterilization containers.

For implants and instruments provided NONSTERILE, sterilization is recommended (wrapped or containerized) as follows:

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Pre-vacuum	132°C (270°F)	4 minutes	30 minutes
Steam	Pre-vacuum	134°C (273°F)	3 minutes	30 minutes

These parameters are validated to sterilize only this device. If other products are added to the sterilizer, the recommended parameters are not valid and new cycle parameters must be established by the user. The sterilizer must be properly installed, maintained, and calibrated. Ongoing testing must be performed to confirm inactivation of all forms of viable microorganisms.